



Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502

Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Group 4 Containers for Inorganics

| <u>Analyte</u> | <u>RL mg/L</u> | <u>Analyte</u> | <u>RL mg/L</u> |
|----------------|----------------|------------------|----------------|
| Ammonia | 0.5 U | TDS | 10 U |
| COD | 15 U | TS | 10 U |
| Nitrate | 0.1 U | TSS | 4 U |
| Nitrite | 0.1 U | Total Phosphorus | 0.05 U |

NOTES:

- a. Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- b. U = The analyte was analyzed for but not detected above the Reporting Limit.
- c. TSS, TDS, and TS analyses do not always contain a QC standard at the Reporting Limit.
- d. Bottles are Type III Soda Lime and vials are Type I Borosilicate.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: 1 mL Hydrochloric Acid (1:1)

Barcoded: Yes

Stirbars: No

Tared Weight: No

Part Number: LPV008611052

Date Product Prepared: 5/12/2025

Item Description: 250cc Natural HDPE WM LEAKPROOF w/
43PP Cap

Lot Number: 051225-4EPJ

Protocol: C Level: 1

Chemical Lot No.: 24015681

Group: 4 (applies)

Chemical Expiry Date: 10/1/2027

 **Chief Executive Officer**