

Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502



Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U.S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Low-Level Polynuclear Aromatic Hydrocarbons (PAHs)

<u>Analyte</u>	<u>MDL ug/L</u>	<u>RL ug/L</u>
1-Methylnaphthalene	0.02U	0.10
2-Methylnaphthalene	0.02U	0.10
Acenaphthene	0.02U	0.10
Acenaphthylene	0.02U	0.10
Anthracene	0.02U	0.10
Benz(a)anthracene	0.02U	0.10
Benzo(a)pyrene	0.02U	0.10
Benzo(b)fluoranthene	0.02U	0.10
Benzo(g,h,i)perylene	0.02U	0.10
Benzo(k)fluoranthene	0.02U	0.10
Chrysene	0.02U	0.10
Dibenz(a,h)anthracene	0.02U	0.10
Fluoranthene	0.02U	0.10
Fluorene	0.02U	0.10
Indeno(1,2,3-cd)pyrene	0.02U	0.10
Naphthalene	0.02U	0.10
Phenanthrene	0.02U	0.10
Pyrene	0.02U	0.10

NOTES:

- a. Method Detection Limit (MDL). Reporting Limit (RL).
- b. U = The analyte was analyzed for but not detected above the MDL.
- c. Bottles are Type III Soda Lime and vials are Type I Borosilicate.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: N/A

Barcoded: Yes

Stirbars: No

Tared Weight: No

Part Number: LW52-A000-A01A-P02

Date Product Prepared: 7/31/2025

Item Description: 1 Liter Amber WM w/ 53-400 White PTFE Lined Cap

Lot Number: 072825-1DDN

Protocol: A Level: 1

Chemical Lot No.: N/A

Group: 1 (applies)

Chemical Expiry Date: N/A

 Chief Executive Officer