



### CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

| <u>Analyte</u> | <u>RL (µmhos/cm)</u> |
|----------------|----------------------|
| Conductivity   | 1.0 U                |

NOTES:

- a. Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- b. U = The analyte was analyzed for but not detected above the Reporting Limit.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Barcoded:  Yes    No    Stirbars:    Yes     No    Tared Weight:    Yes     No

**Level: Quality Assured (QA) (X)**

Glassware / Plasticware received full Quality Assurance and Quality Control treatment. Containers, liners, and closures as applicable, are cleaned according to EPA recommended procedures and validated through a third party (NELAP) testing Laboratory. Each case of containers is custody sealed and labeled for traceability by Lot Number.

Part Number: QLPC15900107

Lot Number: 082525-3BPB

VWR Part No.: 89126-636

Date Product Prepared: 8/27/2025

Item Description: 1 Gallon Cubitainers  
Quality Assured with Labels

Protocol: B    Level: QA

Group: 3 (applies)

Chief Executive Officer  
Manufactured for VWR International