



CERTIFICATE OF ANALYSIS (C OF A)

This "Certificate of Analysis" represents a precleaned product from C & G Containers that has been prepared in accordance with C & G Containers Performance Based criteria and meets or exceeds analyte product specifications per lot number for the analyte, Total Organic Carbon (TOC).

PRODUCTION DATE	PROCESSED CONTROL QA LEVEL	PRODUCT LOT No. / EXPIRATION	ANALYTICAL CERTIFICATION REPORT No.
17 October 2025	Protocol D - 10	Lot#-101725-02EBZ	20OCT2025

PARAMETER ANALYSIS / LIMIT OF DETECTION

ANALYTE	QUANTITY DETECTED
Total Organic Carbon (TOC)	< 25 µg/L / (< 25ppb)

This written document is to certify that a representative Quality Control sample from this lot of containers has been found compliant with C & G Container's specifications for this product.

This completed Certificate is provided for your records and can be used to facilitate any required correspondence as needed.

Quality Assurance / Quality Control Approval:


Lane Breaux – Chief Executive Officer

Date: 21 October 2025

PRODUCT DESCRIPTION/ PRODUCT CONFORMANCE SPECIFICATIONS
Product Number: L-B01-288679

4 oz (120ml) Amber Boston Round bottle. The closure is Polypropylene with a PTFE Liner.

SPECIFICATIONS:
CONTAINER:

CONTAINERS MATERIAL COMPOSITION	PRACTICAL VOLUME	QUANTITY/ CASE	DIMENSIONS (HEIGHT)	DIMENSIONS (WIDTH)
TYPE III - Soda Lime Glass	120 mLs	24	4.406 +/- 0.031	1.906 +/- 0.031

CLOSURE WITH LINER:

CLOSURE MATERIAL COMPOSITION	COLOR	SIZE - FINISH	DIMENSIONS THICKNESS
Polypropylene with PTFE liner	White	22-400	----
Liner	----	22 mm	PTFE – 0.010"

Berlin/Agilent LL Spec.Partic/TOC SOP-050(120 mL/4oz) Amber Boston Round Revision 061923-02

P.O. Box 2003
Lafayette, LA 70502 USA

2202 I-49 N. Service Rd.
Opelousas, La. 70570 USA