

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U.S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Group 2 Containers for Inorganics

<u>Analyte</u>	<u>RL ug/L</u>	<u>Analyte</u>	<u>RL ug/L</u>	<u>Analyte</u>	<u>RL ug/L</u>
Aluminum	30 U	Copper	1 U	Potassium	100 U
Antimony	1 U	Iron	75 U	Selenium	0.5 U
Arsenic	0.5 U	Lead	0.5 U	Silicon	100 U
Barium	0.5 U	Lithium	0.5 U	Silver	0.5 U
Beryllium	0.5 U	Magnesium	30 U	Sodium	50 U
Bismuth	0.5 U	Manganese	0.5 U	Strontium	0.5 U
Boron	10 U	Mercury	0.2 U	Thallium	0.5 U
Cadmium	0.1 U	Molybdenum	0.5 U	Tin	0.5 U
Calcium	100 U	Nickel	0.5 U	Titanium	1 U
Chromium (Total)	2.0 U	Palladium	0.5 U	Vanadium	1 U
Cobalt	1 U	Platinum	0.5 U	Zinc	5 U

<u>Analyte</u>	<u>RL mg/L</u>	<u>Analyte</u>	<u>RL mg/L</u>	<u>Analyte</u>	<u>RL mg/L</u>
Alkalinity	5.0 U	Chloride	0.5 U	Cyanide	0.01 U
Fluoride	0.2 U	Sulfide	0.05 U	Sulfate	0.5 U

NOTE(s):

- a. RL = Reporting Limit
- b. U = The analyte was analyzed for but not detected above the Reporting Limit.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: N/A

Barcoded: No

Stirbars: No

Tared Weight: No

Part Number: LPW066100000

Date Product Prepared: 10/31/2025

Item Description: 2000cc Natural WM w/ 110-400 F217 Lined Cap **Mold Rite Caps**

Lot Number: 102725-2AQO

Protocol: A Level: 1

Chemical Lot No.: N/A

Group: 2 (applies)

Chemical Expiry Date: N/A



Chief Executive Officer