



CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

<u>Analyte</u>	<u>RL (µmhos/cm)</u>
Conductivity	1.0 U

NOTES:

- a. Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- b. U = The analyte was analyzed for but not detected above the Reporting Limit.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Barcoded: Yes No Stirbars: Yes No Tared Weight: Yes No

Level: Processed (PC) (X)

Glassware / Plasticware received full Quality Assurance and Quality Control treatment. Containers, liners, and closures as applicable, are cleaned according to EPA recommended procedures and validated through a third party (NELAP) testing Laboratory. Each case of containers is custody sealed and labeled for traceability by Lot Number.

Part Number: QLPC15900218

Lot Number: 102725-3

VWR Part No.: 89094-140

Date Product Prepared: 10/28/2025

Item Description: 1 Liter LDPE Cubitainer
Processed

Protocol: B Level: PC

Group: 3 (applies)

Chief Executive Officer
Manufactured for VWR International