

PROTOCOL CERTIFICATE WITH INSTRUCTIONS

Enclosed Materials:

1 each Two-piece disposable soil sample (Terra Core™ sampler) 5 Gram 10 Gram

60cc Amber Glass W/M 2oz Clear AC Glass Other

Note: This container may be used for screening and/or % moisture/solids determination.

(2) each	<input type="checkbox"/> 5mLs <input type="checkbox"/> 10mLs <input type="checkbox"/> 15mLs <input type="checkbox"/> Grams <input type="checkbox"/> Other:	<input checked="" type="checkbox"/> No Preservative
Available Preservatives:	Sodium Bisulfate Solution	Stirbars: Yes
Methanol	Sodium Phosphate Tribasic Solution	Barcoded: Yes
Deionized Water	Sodium Phosphate Tribasic Crystal	Tared-Weight: Yes

(1) each	<input type="checkbox"/> 5mLs <input type="checkbox"/> 10mLs <input type="checkbox"/> 15mLs <input type="checkbox"/> Grams <input type="checkbox"/> Other:	<input checked="" type="checkbox"/> No Preservative
Available Preservatives:	Sodium Bisulfate Solution	Stirbars:
Methanol	Sodium Phosphate Tribasic Solution	Barcoded: Yes
Deionized Water	Sodium Phosphate Tribasic Crystal	Tared-Weight: Yes

each	<input type="checkbox"/> 5mLs <input type="checkbox"/> 10mLs <input type="checkbox"/> 15mLs <input type="checkbox"/> Grams <input type="checkbox"/> Other:	<input type="checkbox"/> No Preservative
Available Preservatives:	Sodium Bisulfate Solution	Stirbars:
Methanol	Sodium Phosphate Tribasic Solution	Barcoded:
Deionized Water	Sodium Phosphate Tribasic Crystal	Tared-Weight:

Cleaning Protocol/Lot # and Chemical Traceability # for Glass Only Containers

Item:	40 ml Clear Vial (W)	40 ml Clear Vial (W)	60cc Amber Glass WM
Lot #:	120825-3CLM	120825-3CLM	120825-5
	LKS040000201	LKS040000201	LKS040000201
Protocol/Level:	B/1	B/1	L3
Traceability #:	N/A	N/A	N/A
Chemical Expiry Date:	N/A	N/A	N/A

Recommended Procedural Steps:

1. Have a 40 ml. VOA vial containing the applicable preservative chemical. With the plunger seated in the handle, push the Terra Core™ sampler into the exposed soil type until the chamber is filled. A filled chamber will deliver approximately 5 or 10 grams, but this is dependent on the density of the soil and the syringe size.
2. The soil plug collected should be flush with the open end of the sampler. Wipe away all solids or debris from the outside of the sampler as quickly as possible. (Remove any excess solids that extended beyond the opening of the chamber).
3. Rotate the plunger that was seated in the handle top 90° until it is aligned with the slots in the body. Place the open end of the sampler into the 40-ml vial(s) containing the applicable preservative chemical and slowly extrude the soil plug by pushing the plunger down. (Please note that it may be appropriate to tilt the 40 ml vial as the plug is being delivered to reduce splashing of the preservative chemical). Wipe away any soil or debris from the threads of the vial(s) and quickly place the closure back onto the vial.

Special Notes:

- Each C & G Container's Terra Kit is enclosed in a 2 mil zip top bag.
- The C & G Container's Terra Kit is designed and prepared to hold each initial tared weight measurement as not to exceed a 0.199-gram loss/gain difference from the initial tared weight for six (6) months from the date the kit was prepared.
- The C & G Container Terra Core Kits should be stored in a cool atmosphere and out of direct sunlight.

Should a problem exist or any questions arise, don't hesitate in contacting our technical staff at (800)396-7123.

 **Chief Executive Officer**

Product processed at:
 2202 I-49 N. Service Rd.
 Opelousas LA 70570 USA
 Revision 031020-01sf

Office: 337-237-7123

Toll Free: 800-396-7123

Fax: 337-237-8712

P.O. Box 2003
Lafayette, LA 70502

2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Group 3 Volatile Organic Compounds (VOCs) - Low Level

Analyte	RL ug/Kg	Analyte	RL ug/Kg	Analyte	RL ug/Kg
Acetone	25 U	trans-1,2-Dichloroethylene	1.0 U	n-Propylbenzene	5.0 U
Benzene	0.50 U	1,2-Dichloropropane	1.0 U	Styrene	1.0 U
Bromobenzene	1.0 U	1,3-Dichloropropane	1.0 U	Tert-Amyl Alcohol	10 U
Bromochloromethane	1.0 U	2,2-Dichloropropane	1.0 U	Tert-Amyl Methyl Ether	1.0 U
Bromodichloromethane	1.0 U	1,1-Dichloropropene	1.0 U	Tert-Butyl Alcohol	50 U
Bromoform	1.0 U	cis-1,3-Dichloropropene	1.0 U	1,1,1,2-Tetrachloroethane	1.0 U
n-Butylbenzene	5.0 U	trans-1,3-Dichloropropene	1.0 U	1,1,2,2-Tetrachloroethane	1.0 U
sec-Butylbenzene	5.0 U	Di-Isopropyl Ether	1.0 U	Tetrachloroethylene	1.0 U
tert-Butylbenzene	1.0 U	Ethyl Acetate	10 U	Toluene	5.0 U
Carbon Disulfide	1.0 U	Ethyl Alcohol	400 U	1,2,3-Trichlorobenzene	5.0 U
Carbon Tetrachloride	1.0 U	Ethylbenzene	1.0 U	1,2,4-Trichlorobenzene	5.0 U
Chlorobenzene	1.0 U	Ethyl Tert-Butyl Ether	5.0 U	1,1,1-Trichloroethane	1.0 U
Chloroethane	1.0 U	2-Hexanone	13 U	1,1,2-Trichloroethane	1.0 U
Chloroform	1.0 U	Hexachlorobutadiene	5.0 U	Trichloroethylene	1.0 U
2-Chlorotoluene (ortho)	5.0 U	Isopropyl Alcohol	50 U	Trichlorofluoromethane	1.0 U
4-Chlorotoluene (para)	5.0 U	Isopropylbenzene	5.0 U	1,2,3-Trichloropropane	5.0 U
Dibromochloromethane	1.0 U	p-Isopropyltoluene	5.0 U	1,2,4-Trimethylbenzene	5.0 U
1,2-Dibromo-3-chloropropane	5.0 U	Methyl Acetate	10 U	1,3,5-Trimethylbenzene	5.0 U
1,2-Dibromoethane	1.0 U	Methyl Bromide	10 U	Vinyl Acetate	5.0 U
Dichlorodifluoromethane	1.0 U	Methyl Chloride	5.0 U	Vinyl Chloride	1.0 U
1,3-Dichlorobenzene (meta)	1.0 U	Methylene Bromide	1.0 U	m,p-Xylene	2.0 U
1,2-Dichlorobenzene (ortho)	1.0 U	Methylene Chloride	5.0 U	o-Xylene	1.0 U
1,4-Dichlorobenzene (para)	1.0 U	Methyl Ethyl Ketone	13 U	Xylene (total)	2.0 U
1,1-Dichloroethane	1.0 U	4-Methyl-2-pentanone	13 U		
1,2-Dichloroethane	1.0 U	Methyl Tert Butyl Ether	1.0 U		
1,1-Dichloroethylene	1.0 U	Naphthalene	5.0 U		
cis-1,2-Dichloroethylene	1.0 U				

NOTES:

- Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- U = The analyte was analyzed for but not detected above the Reporting Limit.
- U* = No analytes were detected; No Reporting Limits for these analytes.
- Bottles are Type III Soda Lime and vials are Type I Borosilicate.
- Storage: Store at 85F or 29.4C. Keep away from organic vapors.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Item Description: 40 mL Clear Vials

Lot Number: 120825-3CLM LKS040000201

Protocol: B Level: 1

Group: 3 (applies)

Date Product Prepared: 12/08/25

Product Expiry Date: N/A

 **Chief Executive Officer**