



Office: 337-237-7123  
P.O. Box 2003  
Lafayette, LA 70502

Fax: 337-237-8712  
2202 I-49 N. Service Rd.  
Opelousas, La. 70570

### CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

#### Group 4 Containers for Inorganics

| <u>Analyte</u> | <u>RL mg/L</u> | <u>Analyte</u>   | <u>RL mg/L</u> |
|----------------|----------------|------------------|----------------|
| Ammonia        | 0.5 U          | TDS              | 10 U           |
| COD            | 15 U           | TS               | 10 U           |
| Nitrate        | 0.1 U          | TSS              | 4 U            |
| Nitrite        | 0.1 U          | Total Phosphorus | 0.05 U         |

**NOTES:**

- a. Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- b. U = The analyte was analyzed for but not detected above the Reporting Limit.
- c. TSS, TDS, and TS analyses do not always contain a QC standard at the Reporting Limit.
- d. Bottles are Type III Soda Lime and vials are Type I Borosilicate.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: 2 mL Sulfuric Acid (1:1)

Barcoded: Yes

Stirbars: No

Tared Weight: No

Part Number: LPV016632602

Date Product Prepared: 3/12/2026

Item Description: 500ml Natural HDPE Oblong Leakproof w/  
53PP Cap

Lot Number: 030926-4EKM

Protocol: C Level: 1

Chemical Lot No.: 24017202V

Group: 4 (applies)

Chemical Expiry Date: 11/11/2027

 **Chief Executive Officer**