

Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502



Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a pre-cleaned product from C & G Containers, Inc.'s Pre-clean Division that has been prepared in accordance with EPA Methodology and/or C & G Container's Performance-Based specifications. This product has been tested for Per- and Polyfluoroalkyl Substances (PFAS) by a certified third-party laboratory and determined to be less than the reporting limits (RLs) as indicated below.

Per- and Polyfluoroalkyl Substances (PFAS) in Drinking Water Containers for EPA 537.1

<u>Analyte</u>	<u>RL (ng/L)</u>
Perfluorooctanoic acid (PFOA)	<2
Perfluorooctanesulfonic acid (PFOS)	<2
Perfluorohexanoic acid (PFHxA)	<2
Perfluoroheptanoic acid (PFHpA)	<2
Perfluorononanoic acid (PFNA)	<2
Perfluorodecanoic acid (PFDA)	<2
Perfluoroundecanoic acid (PFUnA)	<2
Perfluorododecanoic acid (PFDoA)	<2
Perfluorotridecanoic acid (PFTrDA)	<2
Perfluorotetradecanoic acid (PFTDA)	<2
Perfluorobutanesulfonic acid (PFBS)	<2
Perfluorohexanesulfonic acid (PFHxS)	<2
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2
Hexafluoropropylene oxide dimer acid (HFPO-DA)	<2
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	<2
9-chlorohexadecafluoro-3-oxanone-1-sulfonic acid (9Cl-PF3ONS)	<2
11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2

NOTES:
Reporting Limit = RL

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Item Description: 250cc Nat HDPE Obl LEAKPROOF 45PP A1/BC/TRIS/TRIS HCL (1.25gms) 24/cs

Barcoded: Yes

Date Product Prepared: 4/2/2026

Part Number: LPV008249602

Chemical Lot No.: 24026729

Lot Number: 010526-2EOX

Chemical Expiry Date: 11/4/2030

Chief Executive Officer