



Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502

Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U. S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Group 4 Containers for Inorganics

<u>Analyte</u>	<u>RL mg/L</u>	<u>Analyte</u>	<u>RL mg/L</u>
Ammonia	0.5 U	TDS	10 U
COD	15 U	TS	10 U
Nitrate	0.1 U	TSS	4 U
Nitrite	0.1 U	Total Phosphorus	0.05 U

NOTES:

- Reporting Limit (RL) = The lowest concentration standard analyzed which can be verified.
- U = The analyte was analyzed for but not detected above the Reporting Limit.
- TSS, TDS, and TS analyses do not always contain a QC standard at the Reporting Limit.
- Bottles are Type III Soda Lime and vials are Type I Borosilicate.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: 2 mL Sulfuric Acid (1:4)

Barcoded: Yes

Stirbars: No

Tared Weight: No

Part Number: LPV008630602S01

Date Product Prepared: 5/18/2026

Item Description: 250cc Natural HDPE Oblong LEAKPROOF
with 45PP Cap

Lot Number: 051826-4EYS

Protocol: C Level: 1

Chemical Lot No.: 24028208

Group: 4 (applies)

Chemical Expiry Date: 1/11/2029

 **Chief Executive Officer**