

Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502



Fax: 337-237-8712
2202 I-49 N. Service Road
Opelousas, LA 70570

CERTIFICATE OF ANALYSIS

This Certificate of Analysis represents a processed product prepared in accordance with C & G Containers' performance based procedures. Representative containers were tested and found to conform to the product specifications listed below.

<u>Test</u>	<u>Specification</u>	<u>Result</u>
Bacterial Plate Count	<0.01 CFU/mL	Pass

NOTES:

- a. CFU = Colony Forming Unit.
- b. PBS = Phosphate Buffered Saline pH 7.4
- c. Storage: Store in clean room bags at room temperature. Keep away from open chemicals and vapors.

This Certificate of Analysis is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: 10 mL of Phosphate Buffered Saline pH 7.4

Part Number: <u>LSTCY0150800</u>	Date Product Prepared: <u>9/07/2026</u>
Item Description: <u>15 mL Sterile Conical Bottom Centrifuge Tube</u>	Lot Number: <u>090726-7EZX</u>
Retest Date: <u>9/07/2027</u>	Chemical Lot No.: <u>M0020260618</u>
	Chemical Expiry Date: <u>6/30/2029</u>

 Chief Executive Officer