Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502

C&G Containers Scientific
"The Total Package!"

Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

# CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U.S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

## Group 2 Containers for Inorganics

| Analyte | RL ug/L | Analyte | RL ug/L | Analyte | RL ug/L |
|---|---|---|---|---|---|
| Aluminum | 30 U | Copper | 1 U | Potassium | 100 U |
| Antimony | 1 U | Iron | 75 U | Selenium | 0.5 U |
| Arsenic | 0.5 U | Lead | 0.5 U | Silicon | 100 U |
| Barium | 0.5 U | Lithium | 0.5 U | Silver | 0.5 U |
| Beryllium | 0.5 U | Magnesium | 30 U | Sodium | 50 U |
| Bismuth | 0.5 U | Manganese | 0.5 U | Strontium | 0.5 U |
| Boron | 10 U | Mercury | 0.2 U | Thallium | 0.5 U |
| Cadmium | 0.1 U | Molybdenum | 0.5 U | Tin | 0.5 U |
| Calcium | 100 U | Nickel | 0.5 U | Titanium | 1 U |
| Chromium (Total) | 2.0 U | Palladium | 0.5 U | Vanadium | 1 U |
| Cobalt | 1 U | Platinum | 0.5 U | Zinc | 5 U |

| Analyte | RL mg/L | Analyte | RL mg/L | Analyte | RL mg/L |
|---|---|---|---|---|---|
| Alkalinity | 5.0 U | Chloride | 0.5 U | Cyanide | 0.01 U |
| Fluoride | 0.2 U | Sulfide | 0.05 U | Sulfate | 0.5 U |
| Ammonia | 0.5 U | TDS | 10 U | Nitrate | 0.1 U |
| COD | 15 U | TS | 10 U | Nitrite | 0.1 U |
| Total Phosphorus | 0.05 U | TSS | 4 U | | |

NOTE(s):
a. RL = Reporting Limit
b. U = The analyte was analyzed for but not detected above the Reporting Limit.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

---

Each container contains: 1 mL Sulfuric Acid (1:1)

*Barcoded:* Yes *Stirbars:* No *Tared Weight:* No

**Part Number:** LPV004631002

**Item Description:** 125cc Natural Leakproof WM w/ 38PP cap

**Protocol:** A/C **Level:** 1

**Group:** 2 (applies)

**Date Product Prepared:** 10/12/2026

**Lot Number:** 101226-4ELR

**Chemical Lot No.:** 24028208

**Chemical Expiry Date:** 1/11/2029

Lane Breaux **Chief Executive Officer**

Product processed at:
2202 I-49 N. Service Rd.
Opelousas LA 70570 USA