Office: 337-237-7123
P.O. Box 2003
Lafayette, LA 70502

C&G Containers Scientific
"The Total Package!"

Fax: 337-237-8712
2202 I-49 N. Service Rd.
Opelousas, La. 70570

CERTIFICATE OF ANALYSIS

This "Certificate of Analysis" represents a precleaned product that has been prepared in accordance with Performance-Based specifications. This product meets or exceeds analyte specifications established in the U.S. EPA OSWER Directive 9240.0-05A "Specification and Guidance for Contaminant-free Sample Containers" for use in Superfund and other Hazardous waste programs.

Group 2 Containers for Inorganics

Analyte	RL ug/L	Analyte	RL ug/L	Analyte	RL ug/L
Aluminum	30 U	Copper	1 U	Potassium	100 U
Antimony	1 U	Iron	75 U	Selenium	0.5 U
Arsenic	0.5 U	Lead	0.5 U	Silicon	100 U
Barium	0.5 U	Lithium	0.5 U	Silver	0.5 U
Beryllium	0.5 U	Magnesium	30 U	Sodium	50 U
Bismuth	0.5 U	Manganese	0.5 U	Strontium	0.5 U
Boron	10 U	Mercury	0.2 U	Thallium	0.5 U
Cadmium	0.1 U	Molybdenum	0.5 U	Tin	0.5 U
Calcium	100 U	Nickel	0.5 U	Titanium	1 U
Chromium (Total)	2.0 U	Palladium	0.5 U	Vanadium	1 U
Cobalt	1 U	Platinum	0.5 U	Zinc	5 U

Analyte	RL mg/L	Analyte	RL mg/L	Analyte	RL mg/L
Alkalinity	5.0 U	Chloride	0.5 U	Cyanide	0.01 U
Fluoride	0.2 U	Sulfide	0.05 U	Sulfate	0.5 U
Ammonia	0.5 U	TDS	10 U	Nitrate	0.1 U
COD	15 U	TS	10 U	Nitrite	0.1 U
Total Phosphorus	0.05 U	TSS	4 U		

NOTE(s):
a. RL = Reporting Limit
b. U = The analyte was analyzed for but not detected above the Reporting Limit.

This "Certificate of Analysis" is provided for your records and is used to facilitate any required correspondences as needed.

Each container contains: N/A

Barcoded: Yes *Stirbars:* No *Tared Weight:* No

Part Number: LPV004200000

Item Description: 125cc Natural HDPE W/M LEAKPROOF w/ 38PP Cap

Protocol: A/C **Level:** 1

Group: 2 (applies)

Date Product Prepared: 11/3/2026

Lot Number: 110226-2ERD

Chemical Lot No.: N/A

Chemical Expiry Date: N/A

Lane Breaux **Chief Executive Officer**

Product processed at:
2202 I-49 N. Service Rd.
Opelousas LA 70570 USA